

HeartLogic™
Heart Failure Diagnostic

SOME TECHNOLOGY IS GAME CHANGING.

THIS IS CAREER DEFINING.

Lead the way with HeartLogic™ from Boston Scientific. The first and only FDA-approved heart failure alert validated to have: high sensitivity, the ability to provide weeks of advanced notice, and low alert burden for detecting worsening heart failure.¹ This is your time. It's your move.

Only available in the Boston Scientific Resonate™ family of CRT-Ds and ICDs.

[LEARN MORE ABOUT HEARTLOGIC](#)

Rx only.

¹ Boehmer JP, Hariharan R, Deveochi FG, et al. A Multisensor algorithm predicts heart failure events in patients with implanted devices: results from the MultiSENSE study. JACC Heart Fail. 2017 Mar;5(3):216-26. CRM-572011-AA

Marek Malik ORCID iD: 0000-0002-3792-1407

Neural Influence of Cardiac Electrophysiology

Marek Malik,¹ Heikki Huikuri,² Federico Lombardi,³ Georg Schmidt,⁴ Markus Zabel⁵
on behalf of the e-Rhythm Study Group of EHRA

1 National Heart and Lung Institute, Imperial College, London, England,

2 Research Unit of Internal Medicine, University Central Hospital and University of Oulu, Finland,

3 Cardiologia, Fondazione IRCCS Ospedale Maggiore Policlinico, Dipartimento di Scienze Cliniche e di Comunità, University of Milan, Italy,

4 Innere Medizin I, Klinikum rechts der Isar, Technische Universität München, München, Germany,

5 Department of Cardiology and Pneumology - Heart Center, University of Göttingen Medical Center, Göttingen, Germany

Address for correspondence:

Marek Malik, PhD, MD, National Heart and Lung Institute, Imperial College,
Dovehouse Street, London SW3 6LY

marek.malik@imperial.ac.uk / marek.malik@btinternet.com

Cardiac electrophysiology is influenced by a spectrum of regulatory mechanisms. Their combined feedback mechanisms influence the electrophysiological processes of every heartbeat to ensure effective response of the cardiovascular system to the demands of the organism. Of these regulatory mechanisms, those maintained at the neural level have been extensively studied but our understanding of the brain – heart interaction is still far from perfect and surely not complete. In particular, while electrophysiologic abnormalities have been reported in patients with different pathologies and abnormalities of the central nervous system,[1-4] only speculations exist on the details of the mechanisms causing these abnormalities.

Some of these central neural and cardiac abnormalities share their basis in cellular electrophysiology. A number of ion channels are important for the function of both

This article has been accepted for publication and undergone full peer review but has not been through the copyediting, typesetting, pagination and proofreading process, which may lead to differences between this version and the Version of Record. Please cite this article as doi: 10.1111/jce.13776.

This article is protected by copyright. All rights reserved.

the brain and the heart [5-7]. Consequently, there are known congenital channelopathies that lead to syndromes with both cardiac and central neural manifestations.

In this issue of the Journal, Hayashi et al report an impressive observational study in patients with epilepsy but without apparent heart disease.[8] Compared with a matched control group, they found more frequent fractionation or slurring of the terminal portion of the QRS complex and, somewhat unexpectedly, marginally albeit statistically significantly shorter QTc interval in the epilepsy patients. This report contributes a valid addition to the mosaic of electrophysiologic abnormalities due to pathologies of the central nervous system. At the same time, since the report by Hayashi et al is a large clinical observational survey, they can only speculate on the underlying mechanisms leading to the observed electrocardiographic differences. The clinical implications of the observed findings are even more limited. Of the two arrhythmic deaths that Hayashi et al observed during the study period, an early repolarization pattern was seen only in one case. It seems therefore seemly to discuss the investigative options that might be proposed for future clinical studies of this kind and that might perhaps advance our understanding of the core mechanistic processes.

Considering the brain-heart interaction, it is not entirely obvious whether QRS fragmentation and the shortened QT interval and other signs of shortened and early repolarization are consistent manifestation of the same process. While early repolarization might share some so-far unknown congenital channelopathy link with epilepsy [9], QRS fragmentation is more likely a reflection of structural and/or functional defects at the level of ventricular myocardium rather than at the level of ionic processes of individual myocytes. It is certainly worthwhile trying to elucidate any links between these cardiac levels and the mechanisms and consequences of the central nervous pathology.

Hayashi et al appropriately relied on the consensus definition of QRS fragmentation and slurring.[10] This consensus definition is binary in the sense of yes/present – no/absent. While such a binary classification is appropriate for clinical assessments, it is not necessarily helpful for detailed physiologic studies since it does not allow any quantification of the degree of the abnormality. Quantifying early repolarization patterns on a continuous numerical scale would be more helpful. We can hope that

interested biomedical engineering groups would soon propose an appropriate methodology so that it can be evaluated in existing ECG collections. Indeed, during the “old days” of signal-averaged electrocardiography, techniques were proposed to quantify abnormalities well within the bulk of the QRS complex.[11,12] While these techniques relied on the high signal to noise ratio due to signal averaging, it seems plausible to adapt their principles to the analysis of individual QRS complexes especially if obtaining the recordings with modern good quality equipment as used by Hayashi et al.

Having a numerical quantifier of early repolarization patterns would be also very helpful in borderline cases. For instance, we note in Figure 3A by Hayashi et al that they observed minimal QRS disturbance (likely well below the thresholds suggested in the consensus document) in some, but not all QRS complexes of lead I while detecting obvious fractionation of lead aVL. It can be debated whether the observation in aVL (which is only a simple algebraic combination of leads I and II that appear to be much less or not at all affected) signifies the same abnormality as in other, say, lateral precordial leads.

We also understand that Hayashi et al evaluated only paper printed electrocardiograms. While this did not compromise their observational study, it would be ideal in future studies if digital recordings were analyzed by objective means independent of visual personal interpretations. Keeping ECG recordings in digital formats of recorded voltage values in individual leads (rather than as only scanned print images) in hospital information systems is now becoming a norm and should be encouraged also for future research purposes. Such recordings could eventually be processed by analytical techniques that are still to come. Studies of heart rate variability have long benefitted from novel processing of historical records of digital tachograms and there are no reasons why morphological analyses of 12-lead ECGs should not follow the same model.

Once morphological ECG abnormalities are quantified on a beat-to-beat basis, the conjecture by Hayashi et al on the autonomic background of the observed abnormalities might also be addressed more directly. Spectral analyses of beat-to-beat fluctuations have long been used to differentiate between the different modulation frequencies that correspond to the feedback control of parasympathetic and

sympathetic regulation. Similarly, numerical quantification of ECG abnormalities would allow studies in patients in whom imaging results of the extent of myocardial scarring and fibrosis are available. This would, among others, not only allow elucidating of whether epileptic seizures lead to histological myocardial abnormalities but also suggest more refined risk factors of sudden unexplained death in epilepsy patients. Assessment of risk in other patient groups would likely benefit equally.

References

- 1 Marafioti V, Turri G, Rossi A, Monaco S, Carbone V. Prolonged QTc interval and insula in patients with ischemic stroke: inductive or abductive reasoning? *Int J Cardiol* 2014; 176:1203-4.
- 2 Bobinger T, Kallmünzer B, Kopp M, Kurka N, Arnold M, Hilz MJ, Huttner HB, Schwab S, Köhrmann M. Prevalence and impact on outcome of electrocardiographic early repolarization patterns among stroke patients: a prospective observational study. *Clin Res Cardiol* 2015; 104:666-71.
- 3 Turri G, Calabrese M, Pancheri E, Monaco S, Gajofatto A, Marafioti V. QTc interval in patients with multiple sclerosis: an inference from the insula of Reil? *Eur J Neurol* 2017; 24:491-6.
- 4 Dabrowski W, Schlegel TT, Wosko J, Rola R, Rzecki Z, Malbrain MLNG, Jaroszynski A. Changes in spatial QRS-T angle and QTc interval in patients with traumatic brain injury with or without intra-abdominal hypertension. *J Electrocardiol* 2018; 51:499-507.
- 5 Simms BA, Souza IA, Zamponi GW. Effect of the Brugada syndrome mutation A39V on calmodulin regulation of Cav1.2 channels. *Mol Brain* 2014; 7:34.
- 6 Andavan GS, Lemmens-Gruber R. Voltage-gated sodium channels: mutations, channelopathies and targets. *Curr Med Chem* 2011; 18:377-97.
- 7 Maljevic S, Wuttke TV, Seeböhm G, Lerche H. KV7 channelopathies. *Pflugers Arch* 2010; 460:277-88.
- 8 Hayashi K, Kohno R, Akamatsu N, Benditt DG, Abe H. *J Cardiovasc Electrophysiol* 2018;29:DOI:10.1111/jce.13746.
- 9 Nolan D, Fink J. Genetics of epilepsy. *Handb Clin Neurol* 2018; 148:467-91.

- 10 Macfarlane PW, Antzelevitch C, Haissaguerre M, Huikuri HV, Potse M, Rosso R, Sacher F, Tikkanen JT, Wellens H, Yan GX. The early repolarization pattern: A consensus paper. *J Am Coll Cardiol* 2015; 66:470-7.
- 11 Ivanov GG, Kovtun VV, Kago M, Titomir LI. High resolution electrocardiographic abnormalities as possible indexes of electric instability of the myocardium. *Can J Cardiol* 1996; 12:53-8.
- 12 Englund A, Andersson M, Bergfeldt L. Spectral turbulence analysis of the signal-averaged electrocardiogram for predicting inducible sustained monomorphic ventricular tachycardia in patients with and without bundle branch block. *Eur Heart J* 1995; 16:1936-42.



ACCURACY MATTERS

SEE THINGS DIFFERENTLY

With a first-of-its-kind, grid-patterned electrode configuration and HD wave bipole recordings along and across the splines, the Advisor™ HD Grid Mapping Catheter, Sensor Enabled™ gives you fast,¹ accurate² high-density cardiac maps. Experience both ease-of-use³ and voltage recordings in direction independent mapping.⁴

Discover fast,¹ accurate² cardiac mapping with the Advisor HD Grid Mapping Catheter, Sensor Enabled.

Learn More at SJM.com/HDGrid

1. Abbott. Report on file. 90299533.
2. Abbott. Report on file. 90262900.
3. Abbott. Report on file. 90355919.
4. Abbott. Report on file. 90280703.

Abbott
One St. Jude Medical Dr., St. Paul, MN 55117 USA
Tel: 1 651 756 2000
SJM.com
St. Jude Medical is now Abbott.

Rx Only

Brief Summary: Prior to using these devices, please review the Instructions for Use for a complete listing of indications, contraindications, warnings, precautions, potential adverse events and directions for use.

™ Indicates a trademark of the Abbott group of companies.
© 2018 Abbott. All Rights Reserved.
28908-SJM-ADV-0818-0042 | Item approved for global use.

